

East Side No 1



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: July 26, 2006 INITIALS: JEC OWNER: WSP  
DISTRICT: 7 ADDRESS: I-5 So. TWP 235  
LOCATION: Bow Hill #33 COUNTY: Skagit  
MAKE: Toledo TYPE: truck scale ACCURACY CLASS: III  
MODEL: 9530 ID. NO: 4409217-4SX CAPACITY: 100000 SECTIONS: 2  
MINIMUM GRADUATIONS: 20 REG. ELEMENT: digital PLATFORM: 12' X 20'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Sec 1</u>	<u>20,000</u>	<u>20000</u>	<u>0</u>				
<u>Center</u>	<u>20 000</u>	<u>20 000</u>	<u>0</u>				
<u>Sec 2</u>	<u>20 000</u>	<u>20020</u>	<u>+20</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. L 2443-1 and 2

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Crona Date: Aug 31, 2006

State of Washington  
County of Thurston  
Signed and sworn to (or affirmed) before  
me on 8-31-06 by

Rita E. Lingle  
(Signature of notary public)

RITA E. LINGLE

Printed Name of Notary

My appointment expires 6-30-10



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS